

April 2021

Attainment of Specific Levels of Relative Humidity Using Saturated Solutions of Chemicals

Mississippi State University

Follow this and additional works at: <https://scholarsjunction.msstate.edu/seedtechpapers>

Recommended Citation

Mississippi State University, "Attainment of Specific Levels of Relative Humidity Using Saturated Solutions of Chemicals" (2021). *Seed Technology Papers*. 22.

<https://scholarsjunction.msstate.edu/seedtechpapers/22>

This Text is brought to you for free and open access by the Mississippi State University Extension Service (MSUES) at Scholars Junction. It has been accepted for inclusion in Seed Technology Papers by an authorized administrator of Scholars Junction. For more information, please contact scholcomm@msstate.libanswers.com.

ATTAINMENT OF SPECIFIC LEVELS OF RELATIVE HUMIDITY USING
SATURATED SOLUTIONS OF THE FOLLOWING
CHEMICALS

<u>SOLID PHASES</u>	<u>°C.</u>	<u>% HUMIDITY</u>
BaCl ₂ ·2H ₂ O -----	24.5	88
CaCl ₂ ·6H ₂ O -----	5.0	39.8
	10.0	38
	18.5	35
	20.0	32.3
	24.5	31
Ca(NO ₃) ₂ ·4H ₂ O - - - - -	18.5	56
	24.5	51
CuSO ₄ ·5H ₂ O -----	20.0	98
CrO ₃ -----	20.0	35
H ₂ C ₂ O ₄ ·2H ₂ O -----	20.0	76
H ₃ PO ₄ ·1/2H ₂ O -----	24.5	9
KC ₂ H ₃ O ₂ -----	20.0	20
	168.0	13
KBr -----	20.0	84
	100.0	69.2
K ₂ CO ₃ ·2H ₂ O -----	18.5	44
	24.5	43
KCNS -----	20.0	47
K ₂ CrO ₄ -----	20.0	88
KF -----	100.0	22.9
K ₂ HPO ₄ -----	20.0	92
KHSO ₄ -----	20.0	86
KI -----	100.0	56.2
KNO ₂ -----	20.0	45
LiCl·H ₂ O -----	20.0	15
MgCl ₂ ·6H ₂ O -----	20.0	35
Mg(C ₂ H ₃ O ₂) ₂ ·4H ₂ O -----	20.0	65
	30.0	60%

<u>SOLID PHASES</u>	<u>°C.</u>	<u>% HUMIDITY</u>
Mg(NO ₃) ₂ ·6H ₂ O	18.5 24.5	56 52
NH ₄ Cl	20.0 25.0 30.0	79.2 79.3 79.5
NH ₄ H ₂ PO ₄	20.0 25.0 30.0	93.1 93.0 92.9
(NH ₄) ₂ SO ₄	20.0 25.0 30.0	81.0 81.1 81.1
NaBr	100.0	22.9
NaBr·2H ₂ O	20.0	58
NaBrO ₃	20.0	92
NaCl	20.0	75
Na ₂ CO ₃ ·10H ₂ O	18.5 24.5	92 87
NaClO ₃	20.0	75
NaC ₂ H ₃ O ₂ ·3H ₂ O	20.0	76
Na ₂ Cr ₂ O ₇ ·2H ₂ O	20.0	52
Na ₂ HPO ₄ ·12H ₂ O	20.0	95
NaHSO ₄ ·H ₂ O	20.0	52
NaNO ₂	20.0	66
Na ₂ SO ₃ ·7H ₂ O	20.0	95
Na ₂ S ₂ O ₃ ·5H ₂ O	20.0	78
Na ₂ SO ₄ ·10H ₂ O	20.0	93
Pb(NO ₃) ₂	20.0	98
ZnCl ₂ ·1 1/2H ₂ O *	20.0	10
→ Zn(NO ₃) ₂ ·6H ₂ O	20.0	42
→ ZnSO ₄ ·7H ₂ O	5.0 20.0	94.7 90.0

* Unstable at this temperature.

% humidity at 20°C. when an excess of the substance indicated is in contact with a saturated aqueous solution of the given solid phase. (From Handbook of Chemistry and Physics.)

<u>Solid phase</u>	<u>% humidity</u>	<u>Solid phase</u>	<u>% humidity</u>
$\text{ZnCl}_2 \cdot 1\frac{1}{2}\text{H}_2\text{O}$	10	$\text{H}_2\text{C}_2\text{O}_4 \cdot 2\text{H}_2\text{O}$	76
$\text{LiCl} \cdot \text{H}_2\text{O}$	15	$\text{Na}_2\text{S}_2\text{O}_3 \cdot 5\text{H}_2\text{O}$	78
$\text{KC}_2\text{H}_3\text{O}_2$	20.	NH_4Cl	79.5
$\text{CaCl}_2 \cdot 6\text{H}_2\text{O}$	32.3	$(\text{NH}_4)_2\text{SO}_4$	81
CrO_3	35	KBr	84
$\text{Zn}(\text{NO}_3)_2 \cdot 6\text{H}_2\text{O}$	42	KHSO_4	86
KNO_2	45	K_2CrO_4	88
KCNS	47	$\text{ZnSO}_4 \cdot 7\text{H}_2\text{O}$	90
$\text{NaHSO}_4 \cdot \text{H}_2\text{O}$	52	NaBrO_3	92
$\text{Na}_2\text{Cr}_2\text{O}_7 \cdot 2\text{H}_2\text{O}$	52	K_2HPO_4	92
$\text{NaBr} \cdot 2\text{H}_2\text{O}$	58	$\text{Na}_2\text{SO}_4 \cdot 10\text{H}_2\text{O}$	93
$\text{Mg}(\text{C}_2\text{H}_3\text{O}_2)_2 \cdot 4\text{H}_2\text{O}$	65	$\text{NH}_4\text{H}_2\text{PO}_4$	93.1
NaNO_2	66	$\text{Na}_2\text{SO}_3 \cdot 7\text{H}_2\text{O}$	95
NH_4Cl and KNO_3	72.6	$\text{Na}_2\text{HPO}_4 \cdot 12\text{H}_2\text{O}$	95
NaClO_3	75	$\text{Pb}(\text{NO}_3)_2$	98
$\text{NaC}_2\text{H}_3\text{O}_2 \cdot 3\text{H}_2\text{O}$	76	$\text{CuSO}_4 \cdot 5\text{H}_2\text{O}$	98

Table I. Glycerine-distilled water solution used to obtain various relative humidity levels at 20° C.

Relative Humidity %	Glycerin %	Water %
20	96	4
42	84	16
48	81	19
52	78	22
60	71	29
70	62	38
75	55	45
78	50	50
88	35	65
96	20	80

The following values were selected from the literature referenced with some values adjusted by interpolation to the temperatures listed.

EQUILIBRIUM RELATIVE HUMIDITIES
FOR SATURATED SALT SOLUTIONS

Saturated Salt Solution	Formula	Percent Relative Humidity at Stated Temperatures		
		68°F(20°C)	77°F(25°C)	86°F(30°C)
Lithium Chloride	$\text{LiCl} \cdot \text{H}_2\text{O}$	12.4	12.0	11.8
Potassium Acetate	$\text{KC}_2\text{H}_3\text{O}_2$	23.3	22.7	22.0
Magnesium Chloride	$\text{MgCl}_2 \cdot 6\text{H}_2\text{O}$	33.6	33.2	32.8
Potassium Carbonate	$\text{K}_2\text{CO}_3 \cdot 2\text{H}_2\text{O}$	44.0	43.8	43.5
Potassium Nitrite	KNO_2	49.0	48.1	47.2
Magnesium Nitrate	$\text{Mg}(\text{NO}_3)_2 \cdot 6\text{H}_2\text{O}$	54.9	53.4	52.0
Sodium Nitrite	NaNO_2	65.3	64.3	63.3
Sodium Chloride	NaCl	75.5	75.8	75.6
Ammonium Sulfate	$(\text{NH}_4)_2\text{SO}_4$	80.6	80.3	80.0
Potassium Nitrate	KNO_3	93.2	92.0	90.7
Potassium Sulfate	K_2SO_4	97.2	96.9	96.6

CONCLUSION

Saturated salt solution generation of humidity is only a secondary method of obtaining an approximate humidity condition. Disagreement among references as to the humidities created indicates a need to qualify this method by independent measurements and to note the salt and reference source used.

To obtain stable humidity values it is essential that the following conditions be met:

1. Chemically pure salt.
2. Distilled water.
3. Large solution surface area and small vapor space.
4. Adequate air circulation.
5. Elimination of hygroscopic construction materials.
6. Solution and vapor at same temperature or in balance.
7. Time allowed for vapor equilibration.

Most of the above requirements (1 thru 5) can be met inexpensively and without elaborate techniques except for the temperature and diffusion rate requirements (6 and 7).

While a difference in the temperatures of the solution and vapor will produce a humidity other than the literature value, a stable humidity condition can be created and maintained without the expense of precise temperature regulation. As long as the temperature relationship between solvent and vapor is kept constant the resultant humidity is also constant. Normally, in a properly designed small chamber the temperatures will be within relatively close agreement so that a stable humidity, within several percent relative humidity of the literature value, will be maintained. However, it is recommended that an independent humidity measuring device, such as HygroDynamics' Hygrosensor, be used to determine the actual humidity condition created and to monitor its stability (4).

The introduction of the hygroscopic material to be tested or conditioned can greatly affect the time required to approach the true equilibrium relative humidity. The mass and hygroscopic nature of the material may be such that fairly stable humidity conditions considerably different from the literature values are created. Therefore, when conditioning or testing a hygroscopic material it is essential that a humidity measuring instrument be used to verify the existing humidity condition.

Saturated salt solution generation of humidity is not as simple as generally believed, but it need not be complex. As noted, the properly designed humidity chamber has adequate air circulation, it minimizes use of hygroscopic materials and temperature unbalance and uses a reliable independent humidity measuring device. For information on chambers meeting these requirements, or if only the humidity measuring equipment is desired, contact HygroDynamics, Inc.